



1.800.299.4142

NO.	DATE	REVISION	BY

TITLE	BBP-ENC-2727-FLXA-FU20-2
DESCRIPTION	DUAL pH SAMPLE SYSTEM
CUSTOMER	

CUSTOMER PO#	DRAWN BY Drew Vaughn
DATE 2/18/2021	CHECKED BY
PROJECT #	REV
DWG #	

PARTS LIST

ITEM	QTY	PART NUMBER	DESCRIPTION
1	2	RAKD41	SS FLOWMETER WITH VALVE
2	2	FF40-S22	FLOWCELL
3	2	PH450 BRACKET	
4	2	K1523JC	FF40 ADAPTER
5	4	FF40 BRACKET	
6	8	92240A578	5/16 SS BOLT
7	8	4830K171	1/2 NPT 316SS CLOSE NIPPLE
8	4	4452K755	1/2 X 1/4 316SS BRANCH TEE
9	4	SS-600-2-4	3/8 TUBE X 1/4" NPT ELBOW
10	2	4452K434	1/2" 316SS TEE
11	2	AA-025-1-F47-3H	0-250F BI-METAL THERMOMETER
12	2	SS-400-1-4	SWAGELOK 1/4 TUBE X 1/4 NPT STRAIGHT
13	2	4003K713	2.5" 0-100PSI SS GAUGE
14	6	SS-8P6T4	1/2" SS PLUG VALVE
15	2	4452K474	1/2" SS STREET ELBOW
16	2	3015T55	VIBRATION-DAMPING ROUTING CLAMP
17	2	SS-600-1-8	3/8 TUBE X 1/2 NPT STRAIGHT CONNECTOR
18	2	3015T42	VIBRATION-DAMPING ROUTING CLAMP
19	2	SS-T6	3/8 STANDOFF
20	2	SS-600-2-8	3/8 TUBE X 1/2 NPT ELBOW
21	2	1443N161	1/2NPT X 3" SS PIPE NIPPLE
22	2	FU20-VS-T1-NPT	PH SENSOR
23	2	FLXA202-D-C-D-CD-C1-NN-A-N-LA-N-NN/UM	SINGLE CHANNEL 2-WIRE ANALYZER
24	2	WU1-V-D-05	5 METER VARIOPIN CABLE
25	1	SCE-36P36SS	33X33 SS SUBPANEL



1.800.299.4142

NO.	DATE	REVISION	BY	TITLE	CUSTOMER PO#	DRAWN BY
						Drew Vaughn
				DESCRIPTION	DATE	CHECKED BY
				DUAL pH SAMPLE SYSTEM	2/18/2021	
				CUSTOMER	PROJECT #	REV
					DWG #	